




SYSTEM CABLES

SOURCE 70101/12
HEAD AMP 70101/15
HEAD AMP+FID 70101/16
GAUGE SI0895
MAINS SI1024
USB SI2288

GC ONLY
70101/16 FIT TO 0-3VFC
70101/15 FIT TO 4-7VFC

SYSTEM CONTROLLER

VALVE BANK 0-----37 D TYPE SOCKET
ON REAR PANEL
PIN 1 TO PIN 1

VALVE BANK 1-----34 WAY IDC TERMINAL
BLOCK

FOR 220-240 VOLTS SET
SWITCH ON SI0508 AS ABOVE

FOR 100-110 VOLTS SET
SWITCH ON SI0508 AS ABOVE

+ TO TOP

34 Way Valve
Terminal Block
SI0373

JH	RELAY CONNECTIONS AMENDED	14/2/2022	C4
DW	MAINS FUSE AND IEC SOCKET ADDED	08/10/19	C3
MOC	TERMINAL BLOCK ADDED	29/4/19	C2
MOC	ELECTRONICS CHANGE	30-5-2014	C1
MOC	ISSUED FOR BUILD	22-12-2009	C
MOC	ORIGINAL	22-12-2009	A
REV.	ISSUED BY	DESCRIPTION	DATE REV. No.
1	REMOVE ALL BURRS & SHARP EDGES	DO NOT SCALE	
MATL	ALL DIMENSIONS IN MILLIMETERS		THIRD ANGLE
FINISH	TOLERANCES UNLESS OTHERWISE STATED		DRG. SIZE A1
	±0.15° ANGULAR	0.00 ±0.1	SCALE NTS
	SURFACE FINISH 1.6μ	0.0 ±0.2	
TITLE	HYDRA 20-22 EDWARDS		DRG No 20402-004-001
	SERCON LTD		

-----VOLTS AC
0.7KVA 50/60 Hz