




SYSTEM CABLES

SOURCE	70101/12
HEAD AMP	70101/15
HEAD AMP+FID	70101/16
GAUGE	SI0895
MAINS	SI1024
USB	SI2288

GC ONLY

70101/16 FIT TO 0-3VFC
70101/15 FIT TO 4-7VFC

SYSTEM CONTROLLER

VALVE BANK 0-----37 D TYPE SOCKET
ON REAR PANEL
PIN 1 TO PIN 1

VALVE BANK 1-----34 WAY IDC TERMINAL
BLOCK

FOR 220-240 VOLTS SET
SWITCH ON SI0508 AS ABOVE

FOR 100-110 VOLTS SET
SWITCH ON SI0508 AS ABOVE

+ TO TOP

34 Way Valve
Terminal Block
SI0373

JH	RELAY CONNECTIONS AMENDED		14/2/2022	C4
DW	MAINS FUSE AND IEC SOCKET ADDED		08/10/19	C3
MOC	TERMINAL BLOCK ADDED		29/4/19	C2
MOC	ELECTRONICS CHANGE		30-5-2014	C1
MOC	ISSUED FOR BUILD		22-12-2009	C
MOC	ORIGINAL		22-12-2009	A
REV.	ISSUED BY	DESCRIPTION	DATE	REV. No.
REMOVE ALL BURRS & SHARP EDGES			DO NOT SCALE	
MATL		ALL DIMENSIONS IN MILLIMETERS	 THIRD ANGLE	
FINISH		TOLERANCES UNLESS OTHERWISE STATED	DRG. SIZE A1	
		±0.15° ANGULAR SURFACE FINISH 1.6μ	0.00 ±0.1 0.0 ±0.2	SCALE NTS
TITLE			DRG No	
HYDRA 20-22 EDWARDS			20402-004-001	
SERCON LTD				
				

-----VOLTS AC

0.7KVA 50/60 Hz