

3 - MACHINE SPECIFICATION

Client:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

Ref No.	Product Description	Required
GC Units <i>continued</i>		
SI2564	Nexis GC-2030	
70901-3	GC Pyrolysis H and O Start Up Kit	
70901-4	GC Pyrolysis H Spares Kit	
70901-5	GC Combustion C Spares Kit (approx. 5000 samples)	
70901-6	GC Pyrolysis h Start Up Kit (approx. 5000 samples)	
71301-1	GC N Spares kit	
ABCA2 Units		
10502/1	ABCA2 Sercon XYZ Autosampler	
10602	ABCA2 Gas Box	
20602	Complete Sercon ABCA2 system	
20603	ABCA IR Analyser	
70101-1	ABCA2 spares kit (Inc. With ABCA2)	
70101-2	ABCA2 Spares Kit for AB Analytica Systems	
70101-3	ABCA2 Autosampler Upgrade Kit	
SI1767	CVL505 - Barcode Reader	
SC9973	13C Urea breath test kit for Helicobacter pylori infection	
SC9974	Helicobacter Test kit for children age 3-11 (45mg)	
Integra2 Units		
20802/4	Integra2	
70201-2	Integra2 analyser spares kit - Included	
70401-1	Integra-2 consumables (Inc. With Integra2)	
Trace Gas Systems		
10802/4	CryoPrep	
10802/5	Automated LN refiller	
10802/6	CryoFlex (TOC)	
10802/7	CryoGas	
10802/8	CryoFlex (LA)	
SI2533	Silicon Carbide Vertical Furnace	
SI2731	213nm Laser Ablation System	
SI2733	isoScell 100 sampler chamber	
71001-1	Cryoprep Spares Kit	
71001-5	CryoGas Consumables Kit	
TOC Units		
14001	Thermalox Environmental TOC-IRMS	
14002	Thermalox Environmental TOC-Standalone	
14003	Thermalox Environmental TOC-Standalone No Sampler	✓
SI2772	CO2 Detector, High sensitivity Option	
SI2773	NOx detector Option	
SI2774	Vacuum pump Option	
SI2775	Stirred Vial rack Option	
72101-1	TOC Start Up Kit	
72101-2	TOC Chem Pack	
72101-3	TOC & TN Instrument with heavy usage - one year	
72101-4	TOC only - economy holding	
GEO 20-22 (dual inlet)		
10702	GEO 20-22 Water Equilibration System (D & 18O)	
21802/1	GEO 20-22	
21802/2	GEO 20-22 + E-Mag	

3 - MACHINE SPECIFICATION

Client:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

21802/3	GEO 20-22 + E-Mag +Hd	
21802/4	GEO 20-22 Inlet	
70301-1	GEO start up kit - Included	
70301-3E	Comprehensive GEO 20-22 spares kit for 5 years operation	

HT-EA

11501/4	HT-EA (With Autosampler)	
11501/5	HT-EA With Liquid Option	
11501/6	HT-EA Stand Alone	
11501/7	HT-EA (Without Autosampler)	
SI2622	CTC PAL3-LSI-L-S	
SI2624	CTC PAL3-Kit-Standalone	
70801-1	Pyrolysis Consumables Kit-High Temp for Solids	
71501-1	HT-EA Pyrolysis Start Up kit O and H	
71501-2	HT-EA Pyrolysis Kit for Water injections	
71501-3	HT-EA Sensitivity Start Up Kit for O & H in small samples	
71501-4	HT-EA Sensitivity Consumables Kit for O & H in small samples	
71501-5	HT-EA Sensitivity Kit for O&H in Small Samples (5000 samples)	
71501-7	HT-EA Kit for Room Temp Equilibration	
71501-8	HT-EA Pyrolysis Kit for Water injections (approx. 1 year kit)	

Ref No.	Product Description	Required
Upgrades		
12101	Evacuation System ***	
12102	2 Ref Gas+ Dilutor Box ***	
20101	Standard Source Assy ***	
20102	HS Source Assy ***	
20301	Standard Collector Assy ***	
30202/1	System Control Upgrade (Pmag) installation not included	Edwards Or Pfeiffer
30202/2	System Controller Upgrade (Emag) installation not included	Edwards Or Pfeiffer
10202/7	Preparation PCB Upgrade	
30202/3	Hijack Box ***	
30302/4	nEXT85 Edwards Pumping System Upgrade Single - installation not included	
30302/5	nEXT85 Edwards Pumping System Upgrade Dual - installation not included	
30302/6	nEXT85 Edwards Pumping System Upgrade Triple- installation not included	
SI2542	ABCA2 Upgrade Kit	
CALLISTOW	Callisto software upgrade ***	
Specials		
11702	LC-I interface	
12202	Leco Interface Unit	
SI2320	Agilent 1260 Series Refractive Index Detection HPLC System	
71401-1	ABCA-G Start Up/Spares Kit	
71401-2	ABCA-G Equilibration Spares Kit - Room temp O only	
71401-3	ABCA-G Equilibration Spares Kit - Room Temp O & H	
71601-1	LC Interface Spares Kit	
ADDITIONS		
PC	Windows 7 based PC System with Callisto Software	
Printer	Printer	
INST1	Installation and start up training at customer's site by Sercon Engineer	
INST2	Installation and start up training at customer's site by Agents Engineer	
WARR	1 year warranty	
WARR2	2 year warranty	
TRAIN	Training in UK for one person for 3 days	
SCONTRACT	Maintenance Service Contract - 1 year	

3 - MACHINE SPECIFICATION


Client:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

SPECIFICATION REQUIREMENTS			
Voltage	Volts 50/60Hzs		
Plug Type	SI2510	3 Metre USA/Japan plug to IEC C 13	
	SI2511	2.5 Metre Australian plug to IEC C 13	
	SI2512	3 Metre UK 13 amp Fused plug to IEC C 13	
	SI2523	3metre Schuko (Euro) Plug to IEC C13 Socket	
	SI2514	2.5 metre Chinese mains to IEC C13	
Pallet	Stamped for export heat treated for 30 min @ 56 °C		
	Complies with all Export Requirements		

3 - MACHINE SPECIFICATION

Client:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

Ref No.	Product Description	Required
HS2022 (continuous flow)		
20402/1	HS2022 Ref Gas + P Mag	
20402/2	HS2022 Ref Gas + E-Mag	
20402/3	HS2022 Ref Gas + Hd Collector + E-Mag	
20402/4	HS2022 Ref Gas + Hd Collector+E-Mag+Trap Loop	
20501	Hd Collector Assy	***
50101/2	Trapping Loop Assy	
70201-1	70201-1 Analyser spares kit – Included	
EA Units		
10202/11	iso-EArth	
10202/13	iso-EArth+	
10202/14	iso-Earth HT Furnace Upgrade	
10202/15	Iso-Earth (Without Autosampler)	
11802/1	Earth+PC	
11802/2	EArth+Clarity+PC	
70501-1	Consumables for C and N for Combustion Only	
70501-11	Simultaneous CNS Consumables Kit (approx. 10000 analysis)	
70501-3	Sample Prep Kit (Inc.)	
70501-4	GSL Combustion and Gas C & N Start Up Kit	
70501-5	Simultaneous NCS Consumables Kit	
70501-6	C&N Combustion Only Consumables Kit (approx. 5000 samples)	
70501-8	Simultaneous CNS Consumables Kit (approx. 5000 samples)	
70601-1	Sulphur Start Up Kit (Inc. with Sercon EA Systems)	
70601-2	Sulphur Consumables Kit for 1000 Samples	
70601-4	Sulphur Consumables Kit (approx. 5000 Samples)	
70801-2	Pyrolysis Consumables Kit-Low temp for Solids	
71701-1	Sub Microgram Nitrogen Analysis Start Up Kit	
SI2040	PN150 Pneumatic Autosampler with one no.50 pos, carousel	
SI2345	ME Micro Balance	
SI2737/A	Sercon 50 Position Autosampler Assy	
SI2789	Extra 50 Position Carousel	
SI2744	Automatic Liquid Sampler 15 Position	
SI2743	Liquid Sampler Mounting Plate	
Headspace Equilibration		
10202/10	XYZ Gas Sampling Control Box	
10302/1	Aqua System	
11103	µCarbs for CF	
11103-1	µCarbs for DI	
11601	200 Place Dual Heated Block 240 Volt	
11801	Sercon XYZ Gas Autosampler	
71107-1	CF oxygen-18 equilibration kit	
71107-2	DI Oxygen-18 Equilibration kit	
71201-1	CF deuterium equilibration kit	
71201-2	DI Deuterium Equilibration kit	
71801-1	WES-DI start up kit (Inc. with WES-DI)	
GC Units		
10402/2	Orchid module 1500C (Will run Hd 70901-1 C Spares Kit Inc)	
10901	Nitrogen Option (71301-1 N Spares kit Included)	
HP7890/1	Agilent 7890B GC System	
HP7890/2	Agilent 7890B GC with Auto Injector	
HP7890/3	Agilent 7890B GC + Auto Injector + 150 Vial Sample Tray	
SI2561	Nexis GC 2010 Plus	

19/01/2017

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

HS20-22			
Reference Gas Precision (n=5)			
¹³ C	(CO ₂)	0.06	
¹⁵ N		0.06	
¹⁸ O	(CO ₂)	0.06	
³⁴ S	(SO ₂)	0.1	
² H		0.4	

Reference Gas Linearity		
¹³ C	≤0.02‰/nA	
¹⁵ N	≤0.02‰/nA	

IsoEarth & IsoEarth+					
Combustion (n=5)					
¹³ C	100 µg C	0.1		10 µg C	0.8
¹⁵ N	100 µg N	0.15		10 µg N	1.5
³⁴ S	25 µg S	0.3		10 µg S	1.0
Dual	400 µg C	0.1		100 µg N	0.15

Integra2			
¹³ C	100 µg C	0.2	
¹⁵ N	100 µg N	0.3	
Dual	100 µg N & 400 µg C	0.4 & 0.15	
³⁴ S	30 µg S	0.5	
¹⁸ O	100 µg O	0.5	

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

HT-EA			
Pyrolysis (n=5)			
¹⁸ O	100 µg O or 1 µl water	0.3	
² H	200 µg H or 1 µl water	1.0	

Aqua, µCarbs & GasBox					
Equilibrations (n=5)					
¹⁸ O	0.2 ml water	0.1			
² H	0.2 ml water	2.0			
Carbonates (n=5)					
¹³ C	100 µg CaCO ₃	0.08			
¹⁸ O	100 µg CaCO ₃	0.1			
Gas Analysis (n=5)					
¹³ C	5% CO ₂	0.1		Air CO ₂	0.1
¹⁸ O	5% CO ₂	0.1		Air CO ₂	0.2
² H	0.5 ml H ₂	1.5			
¹⁵ N	0.5 ml N ₂	0.1		100 ppm N ₂ O	0.5
¹³ C	Breath (12 ml)	0.1			
DIC (n=5)					
¹³ C	15ppm (0.5ml)	0.15			

LC-I			
(n=5)			
¹³ C		0.2	

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

GC Orchid						
Combustion (n=3 for each analyte)						
¹³ C	3 Alkanes	100 ng C	0.2		10 ng C	0.5
¹⁵ N	2 Amines	100 ng N	0.5		14 ng N	1.5
Pyrolysis (n=3 for each analyte)						
² H	3 Alkanes	250 ng H	3.0		25 ng H	10.0
¹⁸ O	3 FAMES	100 ng O	0.5		20 ng O	2.0
Water pyrolysis (n=5)						
² H	0.1 µl water	3.0				
¹⁸ O	0.1 µl water	0.4				

CryoRange						
(n=5)						
¹³ C	Air CO ₂ (12ml)	0.1		Air CH ₄ (125ml)	0.2	
¹⁵ N	Air N ₂ (1ml)	0.1		Air N ₂ O (125ml)	0.4	
¹⁸ O	Air CO ₂ (12ml)	0.15		Air N ₂ O (125ml)	0.7	
¹⁸ O	Air O ₂ (1ml)	0.1				
² H	Air CH ₄	3.0				
¹³ C	TOC (5ppm)	0.4				
¹³ C	Laser					

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

Geo2			
Internal precision (no change overs)			
² H	200 Bar µl	0.15	
¹³ C	100 Bar µl	0.01	
¹⁵ N	100 Bar µl	0.01	
¹⁸ O	100 Bar µl	0.015	
³⁴ S	100 Bar µl	0.015	
Single cold finger volumes (10 change overs)			
¹³ C	20 Bar µl	0.01	
¹⁸ O	20 Bar µl	0.015	
¹³ C	3 Bar µl	0.02	
¹⁸ O	3 Bar µl	0.03	
CAPS (n=5)			
¹³ C	50 µg CaCO ₃	0.03	
¹⁸ O	50 µg CaCO ₃	0.06	
WES (n=5)			
² H		1.0	
¹⁸ O		0.05	
Manifold (10 tubes)			
² H		0.5	
¹³ C		0.02	
¹⁵ N		0.02	
¹⁸ O		0.03	
³⁴ S		0.03	

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

Special tests

Sales Signature:	Date:
Test Signature:	Date:

7: ASSEMBLY AND TEST RECORD

Client:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

PRODUCT:

TOC Standalone w/ CO2 Detector

SERIAL NUMBER:

14002-2/2023/1299

 ASSEMBLED:
 (Name)



Date:

17/5/23

 PRE-TEST
 INSPECTION:
 (Production name)



Date:

17/5/23

 PRE-TEST
 INSPECTION:
 (Test name)

James Ham

Date:

18/5/23

 SYSTEM TEST:
 (Name)

James Ham

Date:

18/5/23

 PRE-DESPATCH
 INSPECTION:
 (Name)

James Ham

Date:

18/5/23

P.A.T. Test results

 In compliance with Low Voltage directive 73/23/EEC
 To Low Voltage standard EN61010 – 1:1993

Earth Continuity

Ω

Insulation

MΩ

Earth Leakage Current

mA. @

kVA.

Tested by:

Date:

8 - Serial Numbers Record Document

Client:	SOTEQ, Chile		
Quotation Number:	16794	Date:	30/11/2022
Job Number:	IN1299	Date:	08/12/2022

Units	S/N
MS/ABCA	
Integra 2	
Prep Unit 1	
Prep Unit 2/Misc	
Electronics Upgrade	
CETAC	
Gas Box	
Autosampler	
BCR	
Dual Heater Block	
TOC 14002-2/2023/1299	
Pump Kit and Optics	S/N
Turbo pump S	
Turbo pump A	
Rotary Pump	
Voltage	110 / 240
Gauge S	
Gauge A	
Source	
Collector 1	
Collector 2	
Magnet	

Parts	S/N
Fan PSU	
24V PSU	

PCBs	S/N
Electronics Box	
Prep PCB	
VFC PCB	
Watchdog PCB	
Head Amp PCB 1	
Head Amp PCB 2	
Electromagnet RP	

Furnaces	S/N
Combustion	
Reduction	
HT-EA	
GC Oven	
Furnace 1	
Furnace 2	

PC	S/N
PC TAG - 3QPQXT3	
Monitor TAG - CGYYFM3	
Printer	

Key

Pumps:	E: Edwards P: Pfeiffer
Furnaces:	V/S: Vecstar E: Elite
Magnet:	E: Electromagnet S: Standard
Turbos & Gauges:	S: Source A: Analyser
Collectors:	Std: Standard S: Sulphur HD: Hydrogen Deuterium

Additional System Info:

CO₂ Detector - 19557
 Syringe Pump - 635465



Packing List

Customer Reference: SPO-469

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Tel: +44 (0)1270 580008
Reference: IN1299

Consignee:

SOTEQ
Mateo de Toro y Zambrano 1491
Of 205, La Reina, Santiago
Chile, 7850329

Contact Name: Leonardo Álvarez
Contact Tel Number: +56948718288

BOX: 1

Quantity	Product Code	Product Description	Serial Number: if available
☒ 1	14002-2	THERMALOX ENVIRONMENTAL TOC STANDALONE CO2	14002-2/2023/1299
☒ 1	SI0155	PC	3QPQXT3
☒ 1	SI2907	MONITOR	CGYYFM3

Packed by:

Date:

James Ham
18/5/23

Claims for damage or shortages must be made within 3 days



S035b_v6

Sercon Limited
Unit 3B Crewe Trade Park
Gateway, Crewe
CW1 6JT U.K.
Company Registration No.: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017



Packing List

Customer Reference: SPO-469

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Tel: +44 (0)1270 580008
Reference: IN1299

Consignee:

SOTEQ
Mateo de Toro y Zambrano 1491
Of 205, La Reina, Santiago
Chile, 7850329

Contact Name: Leonardo Álvarez
Contact Tel Number: +56948718288

BOX: 2

Quantity	Product Code	Product Description	Serial Number: if available
☒ 1	72101-1	TOC START UP KIT	
☒ 2	72101-4	TOC - ECONOMY HOLDING	
☒ 1		VIAL TRAY WITH LID	
☒ 1		FURNACE TOP PLATE	
☒ 1	SI2893	SCRUBBER TUBE	
☒ 2	SI2896	500ML DURAN BOTTLE	
☒ 1		2.5L WASTE BOTTLE	
☒ 3	SI2523	EURO MAINS LEAD	
☒ 1		FURNACE CONNECTOR TUBE	

Packed by:

Date:


18/5/23

Claims for damage or shortages must be made within 3 days



S035b_v6

Sercon Limited
Unit 3B Crewe Trade Park
Gateway, Crewe
CW1 6JT U.K.
Company Registration No.: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017

Packing List

Customer Reference: SPO-469

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Tel: +44 (0)1270 580008
Reference: IN1299

Consignee:

SOTEQ
Mateo de Toro y Zambrano 1491
Of 205, La Reina, Santiago
Chile, 7850329

Contact Name: Leonardo Álvarez
Contact Tel Number: +56948718288

BOX: 1

Quantity	Product Code	Product Description	Serial Number: if available
□		TOC PC MONITOR	

Packed by: _____

Date: _____

Claims for damage or shortages must be made within 3 days

Packing List

Customer Reference: SPO-469

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Tel: +44 (0)1270 580008
Reference: IN1299

Consignee:

SOTEQ
Mateo de Toro y Zambrano 1491
Of 205, La Reina, Santiago
Chile, 7850329

Contact Name: Leonardo Álvarez
Contact Tel Number: +56948718288

BOX: 2

Quantity	Product Code	Product Description	Serial Number: if available
☐	72101-4	lot 2	Economy holding x2
☐	"	2 of 2	"
☐		START UP KIT	
☐		EXTRAS BOX	
☐		- MAINS LEAD X3	
☐		- VIAL RACK + LID	
☐		- FURNACE TOP	
☐		- FURNACE CONNECTOR TUBE	
☐		- 500ML DURAN X2	
☐		- 2.5L WASTE BOTTLE	
☐		- SCRUBBER TUBE.	
☐			

Packed by: _____

Date: _____

Claims for damage or shortages must be made within 3 days



COLLECTION NOTE

Job Number: **IN1299**
Customer: **SOTEQ**
Contact: **Leonardo Alvarez**

Collection on
Date:

23/5/23

No.
of boxes:

2

By:
Company:

NG TRANSPORT

Vehicle
Reg No.

PE13 HVN

Signature:

[Signature]

Print:

C. RICKETTS



S062_v3

Sercon Ltd
Unit 3B Crewe Trade Park
Gateway, Crewe, Cheshire
CW1 6JT U.K.
Registered in England and Wales: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017

Packing List

Customer Reference: SPO-469

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Tel: +44 (0)1270 580008
Reference: IN1299

Consignee:

SOTEQ
Mateo de Toro y Zambrano 1491
Of 205, La Reina, Santiago
Chile, 7850329

Contact Name: Leonardo Álvarez
Contact Tel Number: +56948718288

BOX: Dimensions and weights

	<u>L</u>		<u>W</u>		<u>H</u>			
Box 1	<u>100</u>	x	<u>86</u>	x	<u>110</u>	cm	gw	<u>85.5</u> kg
Box 2	<u>80</u>	x	<u>60</u>	x	<u>62</u>	cm	gw	<u>14</u> kg
Box 3	_____	x	_____	x	_____	cm	gw	_____ kg