

CONTENTS PAGE

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

- ☒ 1 QUOTATION
- ☒ 2 CUSTOMER ORDER
- ☒ 3 MACHINE SPECIFICATION DOCUMENT_S027_v11
- ☒ 4 LOG SHEET_S028_v3
- ☒ 5 FAULT REPORTING DOCUMENT_S029_v3
- ☒ 6 TEST ACCEPTANCE SPECIFICATION_S030_v9
- ☒ 7 ASSEMBLY AND TEST RECORD_S023_v5
- ☒ 8 SERIAL NUMBER FORM_S053_v4
- ☐ 9 TEST DATA
- ☐ 10 PACKING LIST_S035b_v6
- ☐ 11 INSTALLATION REPORT_S057_v6
- ☐ 12 INSTALLATION ACCEPTANCE AND WARRANTY CERTIFICATE_S071_v3

ADDITIONAL:

Payment Terms:

50% / 40% / 10%

Other: please specify

Penalty Clause Yes / No

Please specify: _____

Blue Book Completion:

Signed	<i>Docmor</i>	Date	15/11/23
Verified	<i>J. Webb</i>	Date	15/11/23

3 - MACHINE SPECIFICATION

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

Ref No.	Product Description	Required
HS2022 (continuous flow)		
20402/1	HS2022 Ref Gas + P Mag	
20402/2	HS2022 Ref Gas + E-Mag	
20402/3	HS2022 Ref Gas + Hd Collector + E-Mag	
20402/4	HS2022 Ref Gas + Hd Collector+E-Mag+Trap Loop	✓
20501	Hd Collector Assy	***
50101/2	Trapping Loop Assy	
70201-1	70201-1 Analyser spares kit – Included	✓
EA Units		
10202/11	iso-EArth	
10202/13	iso-EArth+	✓
10202/14	iso-Earth HT Furnace Upgrade	
10202/15	Iso-Earth (Without Autosampler)	
11802/1	Earth+PC	
11802/2	EArth+Clarity+PC	
70501-1	Consumables for C and N for Combustion Only	
70501-11	Simultaneous CNS Consumables Kit (approx. 10000 analysis)	
70501-3	Sample Prep Kit (Inc.)	
70501-4	GSL Combustion and Gas C & N Start Up Kit	
70501-5	Simultaneous NCS Consumables Kit	
70501-6	C&N Combustion Only Consumables Kit (approx. 5000 samples)	
70501-8	Simultaneous CNS Consumables Kit (approx. 5000 samples)	
70601-1	Sulphur Start Up Kit (Inc. with Sercon EA Systems)	
70601-2	Sulphur Consumables Kit for 1000 Samples	
70601-4	Sulphur Consumables Kit (approx. 5000 Samples)	
70801-2	Pyrolysis Consumables Kit-Low temp for Solids	
71701-1	Sub Microgram Nitrogen Analysis Start Up Kit	
SI2040	PN150 Pneumatic Autosampler with one no.50 pos, carousel	
SI2345	ME Micro Balance	
SI2737/A	Sercon 50 Position Autosampler Assy	
SI2789	Extra 50 Position Carousel	
SI2744	Automatic Liquid Sampler 15 Position	
SI2743	Liquid Sampler Mounting Plate	
Headspace Equilibration		
10202/10	XYZ Gas Sampling Control Box	
10302/1	Aqua System	
11103	µCarbs for CF	
11103-1	µCarbs for DI	
11601	200 Place Dual Heated Block 240 Volt	
11801	Sercon XYZ Gas Autosampler	
71107-1	CF oxygen-18 equilibration kit	
71107-2	DI Oxygen-18 Equilibration kit	
71201-1	CF deuterium equilibration kit	
71201-2	DI Deuterium Equilibration kit	
71801-1	WES-DI start up kit (Inc. with WES-DI)	
GC Units		
10402/2	Orchid module 1500C (Will run Hd 70901-1 C Spares Kit Inc)	
10901	Nitrogen Option (71301-1 N Spares kit Included)	
HP7890/1	Agilent 7890B GC System	
HP7890/2	Agilent 7890B GC with Auto Injector	
HP7890/3	Agilent 7890B GC + Auto Injector + 150 Vial Sample Tray	
SI2561	Nexis GC 2010 Plus	

3 - MACHINE SPECIFICATION

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

Ref No.	Product Description	Required
GC Units <i>continued</i>		
SI2564	Nexis GC-2030	
70901-3	GC Pyrolysis H and O Start Up Kit	
70901-4	GC Pyrolysis H Spares Kit	
70901-5	GC Combustion C Spares Kit (approx. 5000 samples)	
70901-6	GC Pyrolysis h Start Up Kit (approx. 5000 samples)	
71301-1	GC N Spares kit	
ABCA2 Units		
10502/1	ABCA2 Sercon XYZ Autosampler	
10602	ABCA2 Gas Box	
20602	Complete Sercon ABCA2 system	
20603	ABCA IR Analyser	
70101-1	ABCA2 spares kit (Inc. With ABCA2)	
70101-2	ABCA2 Spares Kit for AB Analytica Systems	
70101-3	ABCA2 Autosampler Upgrade Kit	
SI1767	CVL505 – Barcode Reader	
SC9973	13C Urea breath test kit for Helicobacter pylori infection	
SC9974	Helicobacter Test kit for children age 3-11 (45mg)	
Integra2 Units		
20802/4	Integra2	
70201-2	Integra2 analyser spares kit - Included	
70401-1	Integra-2 consumables (Inc. With Integra2)	
Trace Gas Systems		
10802/4	CryoPrep	
10802/5	Automated LN refiller	
10802/6	CryoFlex (TOC)	
10802/7	CryoGas	
10802/8	CryoFlex (LA)	
SI2533	Silicon Carbide Vertical Furnace	
SI2731	213nm Laser Ablation System	
SI2733	isoScell 100 sampler chamber	
71001-1	Cryoprep Spares Kit	
71001-5	CryoGas Consumables Kit	
TOC Units		
14001	Thermalox Environmental TOC-IRMS	
14002	Thermalox Environmental TOC-Standalone	
14003	Thermalox Environmental TOC-Standalone No Sampler	
SI2772	CO2 Detector, High sensitivity Option	
SI2773	NOx detector Option	
SI2774	Vacuum pump Option	
SI2775	Stirred Vial rack Option	
72101-1	TOC Start Up Kit	
72101-2	TOC Chem Pack	
72101-3	TOC & TN Instrument with heavy usage - one year	
72101-4	TOC only - economy holding	
GEO 20-22 (dual inlet)		
10702	GEO 20-22 Water Equilibration System (D & 18O)	
21802/1	GEO 20-22	
21802/2	GEO 20-22 + E-Mag	

3 - MACHINE SPECIFICATION

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022


21802/3	GEO 20-22 + E-Mag +Hd	
21802/4	GEO 20-22 Inlet	
70301-1	GEO start up kit - Included	
70301-3E	Comprehensive GEO 20-22 spares kit for 5 years operation	
HT-EA		
11501/4	HT-EA (With Autosampler)	✓
11501/5	HT-EA With Liquid Option	
11501/6	HT-EA Stand Alone	
11501/7	HT-EA (Without Autosampler)	
SI2622	CTC PAL3-LSI-L-S	
SI2624	CTC PAL3-Kit-Standalone	
70801-1	Pyrolysis Consumables Kit-High Temp for Solids	✓
71501-1	HT-EA Pyrolysis Start Up kit O and H	
71501-2	HT-EA Pyrolysis Kit for Water injections	
71501-3	HT-EA Sensitivity Start Up Kit for O & H in small samples	✓
71501-4	HT-EA Sensitivity Consumables Kit for O & H in small samples	
71501-5	HT-EA Sensitivity Kit for O&H in Small Samples (5000 samples)	✓
71501-7	HT-EA Kit for Room Temp Equilibration	
71501-8	HT-EA Pyrolysis Kit for Water injections (approx. 1 year kit)	

Ref No.	Product Description	Required
Upgrades		
12101	Evacuation System ***	
12102	2 Ref Gas+ Dilutor Box ***	
20101	Standard Source Assy ***	
20102	HS Source Assy ***	
20301	Standard Collector Assy ***	
30202/1	System Control Upgrade (Pmag) installation not included	Edwards Or Pfeiffer
30202/2	System Controller Upgrade (Emag) installation not included	Edwards Or Pfeiffer
10202/7	Preparation PCB Upgrade	
30202/3	Hijack Box ***	
30302/4	nEXT85 Edwards Pumping System Upgrade Single - installation not included	
30302/5	nEXT85 Edwards Pumping System Upgrade Dual - installation not included	
30302/6	nEXT85 Edwards Pumping System Upgrade Triple- installation not included	
SI2542	ABCA2 Upgrade Kit	
CALLISTOW	Callisto software upgrade ***	
Specials		
11702	LC-I interface	
12202	Leco Interface Unit	
SI2320	Agilent 1260 Series Refractive Index Detection HPLC System	
71401-1	ABCA-G Start Up/Spares Kit	
71401-2	ABCA-G Equilibration Spares Kit - Room temp O only	
71401-3	ABCA-G Equilibration Spares Kit - Room Temp O & H	
71601-1	LC Interface Spares Kit	
ADDITIONS		
PC	Windows 7 based PC System with Callisto Software	
Printer	Printer	
INST1	Installation and start up training at customer's site by Sercon Engineer	
INST2	Installation and start up training at customer's site by Agents Engineer	
WARR	1 year warranty	
WARR2	2 year warranty	
TRAIN	Training in UK for one person for 3 days	
SCONTRACT	Maintenance Service Contract – 1 year	

3 - MACHINE SPECIFICATION


Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022


SPECIFICATION REQUIREMENTS		
Voltage	Volts 50/60Hzs	
Plug Type	SI2510	3 Metre USA/Japan plug to IEC C 13
	SI2511	2.5 Metre Australian plug to IEC C 13
	SI2512	3 Metre UK 13 amp Fused plug to IEC C 13
	SI2523	3metre Schuko (Euro) Plug to IEC C13 Socket
	SI2514	2.5 metre Chinese mains to IEC C13
Pallet	Stamped for export heat treated for 30 min @ 56 °C	
	Complies with all Export Requirements	



Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

[illegible]





Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

[illegible]

4 - LOG SHEET

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

Date	Action Taken	By Whom
29 MAR 23	Set up 20-22. Comm OK Setup software for Iso Earth + HTGA - All OK. Pumped down + conditioned filament.	DH
30 MAR 23	Set up plumbing + tubes for Iso Earth Furnace set up. Heat up OK. Adjusted HTGA board jumpers to need HT furnace thermocouple type. Poor peak shapes - unstable beam. Replaced emag. Now stable. Still has poor peak shapes. Replaced source with Test Plate. Peaks + Sens much better. New source to be built + fitted	DH
31 MAR 23	Tested Stability + Prec + Linearity on Test Rig Same. All OK Filament Unit on N ₂ tuning	DH
31 MAR 23	Removed Head Amp to test IN1291 Part to be returned.	DH
5 April	Source unstable on CO ₂ settings as the test source was switched for the original with a "NEW" filament. Pumped down over night. Tubes removed & checked in EA. Auto-sampler torqued and setup up ready for testing. Head amp reinstalled - short tube source housing badly contaminated. Cleaned off and popping for the best. Conditioned new filament	SB
5 April		SB

4 - LOG SHEET

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

Date	Action Taken	By Whom
11-4-23	Source turning up with unusual settings -44 Vp & +11 repeler and sensitivity V poor. The above gave terrible peak shapes that could not be improved. Source removed and I found that one ceramic pillar had blocked this was replaced and the system switched to pump down. Source still very insensitive Source swapped for test source	SB LM
12-4-23	Test source runs very well in this system. Showing problem is with the source. Source replaced with test source then changed out for a replacement source. Could not fit the new source has the locating pins on the source will not fit into the locating holes. Source badly damaged after trying to remove it from the flight tube.	SB
17-4-23	Harry rebuilt source after the locating holes had been received out by Mark.	SB, HR, MC
18-4-23	Peak shapes still will not improve regardless of tuning or magnification Swapped out source for test rig source and left instrument to pump down.	SB
19-4-23	Rebuilt original source with York source block and installed this into instrument	SB
21 Apr 23	Filament conditioned away night before Emissions look good at 90 trap @ 50.	DM

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

GC Orchid						
Combustion (n=3 for each analyte)						
¹³ C	3 Alkanes	100 ng C	0.2		10 ng C	0.5
¹⁵ N	2 Amines	100 ng N	0.5		14 ng N	1.5
Pyrolysis (n=3 for each analyte)						
² H	3 Alkanes	250 ng H	3.0		25 ng H	10.0
¹⁸ O	3 FAMES	100 ng O	0.5		20 ng O	2.0
Water pyrolysis (n=5)						
² H	0.1 µl water	3.0				
¹⁸ O	0.1 µl water	0.4				

CryoRange						
(n=5)						
¹³ C	Air CO ₂ (12ml)	0.1		Air CH ₄ (125ml)	0.2	
¹⁵ N	Air N ₂ (1ml)	0.1		Air N ₂ O (125ml)	0.4	
¹⁸ O	Air CO ₂ (12ml)	0.15		Air N ₂ O (125ml)	0.7	
¹⁸ O	Air O ₂ (1ml)	0.1				
² H	Air CH ₄	3.0				
¹³ C	TOC (5ppm)	0.4				
¹³ C	Laser					

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

Geo2			
Internal precision (no change overs)			
² H	200 Bar µl	0.15	
¹³ C	100 Bar µl	0.01	
¹⁵ N	100 Bar µl	0.01	
¹⁸ O	100 Bar µl	0.015	
³⁴ S	100 Bar µl	0.015	
Single cold finger volumes (10 change overs)			
¹³ C	20 Bar µl	0.01	
¹⁸ O	20 Bar µl	0.015	
¹³ C	3 Bar µl	0.02	
¹⁸ O	3 Bar µl	0.03	
CAPS (n=5)			
¹³ C	50 µg CaCO ₃	0.03	
¹⁸ O	50 µg CaCO ₃	0.06	
WES (n=5)			
² H		1.0	
¹⁸ O		0.05	
Manifold (10 tubes)			
² H		0.5	
¹³ C		0.02	
¹⁵ N		0.02	
¹⁸ O		0.03	
³⁴ S		0.03	



Test acceptance - 6

Sercon Ltd, Unit 3B Crewe Trade Park, Gateway
Crewe, Cheshire CW1 6JT, UK
Tel: +44 (0)1270 580008 Fax: +44 (0)1270 252310
Company Registration No. 4290072 Vat No. 771 1100 74

6 - TEST / ACCEPTANCE SPECIFICATION

Customer:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

Special tests

Sales Signature: <u>Go'canner</u>	Date: <u>5/5/23</u>
Test Signature: <u>S. Madley</u>	Date: <u>5/5/23</u>

7: ASSEMBLY AND TEST RECORD

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

PRODUCT:

HS2022

SERIAL NUMBER:

20402/4-2023/1295

 ASSEMBLED:
 (Name)

[Signature]

Date:

10/3/23

 PRE-TEST
 INSPECTION:
 (Production name)

P. Mayo.

Date:

10/3/23

 PRE-TEST
 INSPECTION:
 (Test name)

S. Moss

Date:

10/3/23

 SYSTEM TEST:
 (Name)

S. Moss

Date:

27/4/23

 PRE-DESPATCH
 INSPECTION:
 (Name)

[Signature] [Signature]

Date:

04/05/23

P.A.T. Test results

 In compliance with Low Voltage directive 73/23/EEC
 To Low Voltage standard EN61010 – 1:1993

Earth Continuity

0.075

Ω

Insulation

7300

MΩ

Earth Leakage Current

0.60

mA. @

0.05

kVA.

Tested by:

P.H

Date:

10/3/23

7: ASSEMBLY AND TEST RECORD

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

PRODUCT:

HT-EA (WITH AUTO)

SERIAL NUMBER:

11501/4-2023/1295

 ASSEMBLED:
 (Name)

[Signature]

Date:

3/4/23

 PRE-TEST
 INSPECTION:
 (Production name)

P. Hayes

Date:

3/4/23

 PRE-TEST
 INSPECTION:
 (Test name)

S. Moses

Date:

3/4/23

 SYSTEM TEST:
 (Name)

S. Moses

Date:

27/4/23

 PRE-DESPATCH
 INSPECTION:
 (Name)

[Signature] [Signature]

Date:

02/05/23

P.A.T. Test results

 In compliance with Low Voltage directive 73/23/EEC
 To Low Voltage standard EN61010 – 1:1993

Earth Continuity

Ω

Insulation

MΩ

Earth Leakage Current

mA. @ _____ kVA.

Tested by:

Date:

7: ASSEMBLY AND TEST RECORD

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

PRODUCT:

ISO-EARTH+

SERIAL NUMBER:

10202/13-2023/
15281.

 ASSEMBLED:
 (Name)

GOODS RETURNED

Date:

 PRE-TEST
 INSPECTION:
 (Production name)

SEE IN1254
GULF BIO
=

Date:

 PRE-TEST
 INSPECTION:
 (Test name)

Date:

 SYSTEM TEST:
 (Name)

Date:

 PRE-DESPATCH
 INSPECTION:
 (Name)

Date:

P.A.T. Test results

 In compliance with Low Voltage directive 73/23/EEC
 To Low Voltage standard EN61010 – 1:1993

Earth Continuity

Ω

Insulation

MΩ

Earth Leakage Current

mA. @ _____ kVA.

Tested by:

Date:

8 - Serial Numbers Record Document

Client:	Ewha Womans Univ. C/O: Saram International Co. Ltd		
Quotation Number:	14556	Date:	16/11/2022
Job Number:	IN1295	Date:	24/11/2022

Units	S/N
MSI/ABCA	20402/4-2023/1295
Integra 2 /	
Prep Unit 1	11501/4-2023/1295
Prep Unit 2/Misc	10202/13-2023/1295
Electronics Upgrade /	
CETAC /	
Gas Box /	
Autosampler	EARTH 42670 HTEA 42671
BCR /	
Dual Heater Block /	

Pump Kit and Optics	S/N
Turbo pump S	220791826
Turbo pump A	220786639
Rotary Pump	220510672
Voltage	110 / (240)
Gauge S	220791826 ^{HR} 220541900
Gauge A	220786639 ^{HR} 086205223
Source	20102MS-210
Collector 1	20301-135
Collector 2	20501-053
Magnet /	

Parts	S/N
Fan PSU /	
24V PSU /	

PCBs	S/N
Electronics Box	20230008
Prep PCB	EARTH 20200002 HTEA 20230001
VFC PCB	23005
Watchdog PCB	
Head Amp PCB 1	21003
Head Amp PCB 2	22017
Electromagnet RP	133

Furnaces	S/N
Combustion /	
Reduction /	
HT-EA /	
GC Oven /	
Furnace 1 /	
Furnace 2 /	

PC	S/N
PC	GC5C7W3
Monitor	GRDJ3H3
Printer /	

Key

Pumps:	E: Edwards P: Pfeiffer
Furnaces:	V/S: Vecstar E: Elite
Magnet:	E: Electromagnet S: Standard
Turbos & Gauges:	S: Source A: Analyser
Collectors:	Std: Standard S: Sulphur HD: Hydrogen Deuterium

Additional System Info:



Packing List

Customer Reference: SR2022-P1124-E001_EWHA

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Consignee/Importer:

Saram International Co., Ltd
B1019, F10, 167
Songpa-daero, Songpa-gu
Seoul, 05855, Korea
Contact Name: June Yoo
Tel Number: +82 2 3402 0518

Delivery Address:

EWHA WOMAN'S UNIVERSITY
B-356, Education Building
52, Ewhayeodae-gil, Seodaemun-gu, Seoul
03760
South Korea

Tel: +44 (0)1270 580008

Reference: IN1295

BOX: 1

Quantity	Product Code	Product Description	Serial Number: if available
☒ 1	20402	HS2022	20402/4-2023/1295

Packed by:

Date:

02/05/2023

Claims for damage or shortages must be made within 3 days



S035b_v6

Sercon Limited
Unit 3B Crewe Trade Park
Gateway, Crewe
CW1 6JT U.K.
Company Registration No.: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017



Packing List

Customer Reference: SR2022-P1124-E001_EWHA

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Consignee/Importer:

Saram International Co., Ltd
B1019, F10, 167
Songpa-daero, Songpa-gu
Seoul, 05855, Korea
Contact Name: June Yoo
Tel Number: +82 2 3402 0518

Delivery Address:

EWHA WOMAN'S UNIVERSITY
B-356, Education Building
52, Ewhayeodae-gil, Seodaemun-gu, Seoul
03760
South Korea

Tel: +44 (0)1270 580008

Reference: IN1295

BOX: 2

Quantity	Product Code	Product Description	Serial Number: if available
☒ 1	10202	ISO-EARTH	10202/13-2023/1295

Packed by:

Date:



02/05/2023

Claims for damage or shortages must be made within 3 days



S035b_v6

Sercon Limited
Unit 3B Crewe Trade Park
Gateway, Crewe
CW1 6JT U.K.
Company Registration No.: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017



Packing List

Customer Reference: SR2022-P1124-E001_EWHA

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Consignee/Importer:

Saram International Co., Ltd
B1019, F10, 167
Songpa-daero, Songpa-gu
Seoul, 05855, Korea
Contact Name: June Yoo
Tel Number: +82 2 3402 0518

Delivery Address:

EWHA WOMAN'S UNIVERSITY
B-356, Education Building
52, Ewhayeodae-gil, Seodaemun-gu, Seoul
03760
South Korea

Tel: +44 (0)1270 580008
Reference: IN1295

BOX: 3

Quantity	Product Code	Product Description	Serial Number: if available
☒ 1	11501	HT-EA	11501/4-2023/1295

Packed by:

Date:


02/05/2023

Claims for damage or shortages must be made within 3 days



S035b_v6

Sercon Limited
Unit 3B Crewe Trade Park
Gateway, Crewe
CW1 6JT U.K.
Company Registration No.: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017



Packing List

Customer Reference: SR2022-P1124-E001_EWHA

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Consignee/Importer:

Saram International Co., Ltd
B1019, F10, 167
Songpa-daero, Songpa-gu
Seoul, 05855, Korea
Contact Name: June Yoo
Tel Number: +82 2 3402 0518

Delivery Address:

EWHA WOMAN'S UNIVERSITY
B-356, Education Building
52, Ewhayeodae-gil, Seodaemun-gu, Seoul
03760
South Korea

Tel: +44 (0)1270 580008

Reference: IN1295

BOX: 4

Quantity	Product Code	Product Description	Serial Number: if available
☒	1	SI2391	PC
☒	1	SI2391	Monitor
☒	1		Software backup USB (inside pc box)
☒	1		Windows recovery USB (inside pc box)
☒	1	SI0103	Rotary pump Edwards
☒	1	SI0685	Electro magnet
☒	2	SI2737	Autosampler
☒	1		External cable and pipe box
☒	2		Euro mains cable (inside external cable and pipes box)
☒	1		External pipes (inside external cable and pipes box)
☒	1		HTEA 16 amp cable
☒	1	70201-1	Analyser spare kit
☒	1	70801-1	Pyrolysis start up kit
☒	1	71501-1	HT-EA pyrolysis start up kit
☒	1	70501-3	Sample preparation kit
☒	1	70501-5	Start-up consumables kit

Packed by:

Date:

02/05/2023

Claims for damage or shortages must be made within 3 days



S035b_v6

Sercon Limited
Unit 3B Crewe Trade Park
Gateway, Crewe
CW1 6JT U.K.
Company Registration No.: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017

Packing List

Customer Reference: SR2022-P1124-E001_EWHA

Shipper / Exporter:

Sercon Limited
Unit 3B Crewe Trade park
Gateway
Crewe
CW1 6JT
UK

Consignee/Importer:

Saram International Co., Ltd
B1019, F10, 167
Songpa-daero, Songpa-gu
Seoul, 05855, Korea
Contact Name: June Yoo
Tel Number: +82 2 3402 0518

Delivery Address:

EWHA WOMAN'S UNIVERSITY
B-356, Education Building
52, Ewhayeodae-gil, Seodaemun-gu, Seoul
03760
South Korea

Tel: +44 (0)1270 580008

Reference: IN1295

BOX: Dimensions and weights

	<u>L</u>		<u>W</u>		<u>H</u>		
Box 1	99	x	85	x	107	cm	gw 135 kg
Box 2	99	x	85	x	90	cm	gw 108 kg
Box 3	99	x	85	x	80	cm	gw 89 Kg
Box 4	99	x	85	x	107	cm	gw 99 kg



COLLECTION NOTE

Job Number: **IN1295**
Customer: **Ewha Womans Univ. C/O: Saram International Co. Ltd**
Contact: **June Yoo**

Collection on

Date:

3-5-23

No.

of boxes:

4

By:

Company:

NG TRANS

Vehicle

Reg No.

~~AB~~ MT22DZN

Signature:

[Signature]

Print:

P. THE



S062_v3

Sercon Ltd
Unit 3B Crewe Trade Park
Gateway, Crewe, Cheshire
CW1 6JT U.K.
Registered in England and Wales: 4290072

Phone: +44 (0) 1270 580008
Fax: +44 (0) 1270 252310
sales@sercongroup.com
www.sercongroup.com
VAT No.: 771 1100 74
19/01/2017