

4 - LOG SHEET

Client:	San Francisco State University		
Quotation Number:	11414	Date:	11/10/2021
Job Number:	IN1272	Date:	04/01/2022

Date	Action Taken	By Whom
25 Jan 21	filled Electronics upgrade system + cables to Test Rig 2. Comms OK to PC Pumped down instrument. All read backs respond OK. Fired filament - all OK. Pump down over night.	DJH 25 Jan 21
26 Jan 21	Beams respond OK. CO₂ peak shape + stability - OK N₂ peak shape + stability - OK Ran CO₂ + N₂ prec CO₂ = 0.024%. N₂ = 0.063%. No issues seen.	DJH 26 Jan 21
27 Jan 21	Re-checked beams. All OK. System good to ship. Parts removed from Test Rig 2. Need to add Sys Con to housing + test cooling fans.	DJH