



Sercon Ltd, Unit 3B Crewe Trade Park, Gateway
Crewe, Cheshire CW1 6JT, UK
Tel: +44 (0)1270 580008 Fax: +44 (0)1270 252310
Company Registration No. 4290072 Vat No. 771 1100 74

6: SYSTEM CONTROLLER TEST REPORT

Client:	San Francisco State University		
Quotation Number:	11414	Date:	11/10/2021
Job Number:	IN1272	Date:	04/01/2022

Serial Number:	70220001	Instrument Configuration:	
Tested By:	D. HARRIS	Date:	27 Jan 22

Initial Tests			
Initial communication with PC	Initials: DJH	Date:	27 Jan 22
Turbo speed readbacks respond	Initials: DJH	Date:	27 Jan 22
Pressure trip limits are set and respond	Initials: DJH	Date:	27 Jan 22
Vacuum gauges readbacks respond to changes in pressure	Initials: DJH	Date:	27 Jan 22
Filament switches on and regulates normally	Initials: DJH	Date:	27 Jan 22
All source readbacks match and respond to changes in set values	Initials: DJH	Date:	27 Jan 22
E Magnet limit is set and readbacks follow set values (if appropriate)	Initials: N/A	Date:	N/A
Valve banks respond (ensure all appropriate banks are checked)	Initials: DJH	Date:	27 Jan 22
Flow meters respond appropriately	Initials: N/A	Date:	N/A
Gas pressure gauges respond appropriately	Initials: N/A	Date:	N/A
N ₂ and CO ₂ peaks and stability are good	Initials: DJH	Date:	27 Jan 22
HD system VFC fitted and tested (if appropriate)	Initials: N/A	Date:	N/A
All cables, boxes and equipment have been packed	Initials:	Date:	

System Controller Sign Off			
<i>Ensure all boxes are filled in, signed and Dated as appropriate. If a test is not applicable the mark the associated boxes with N/A (not applicable)</i>			
Tests Completed By:		Date:	27 Jan 21
Dispatched By:	Go' Connor	Date:	8/2/22